

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 11 of 11

Proficiency Test Report

Lab Id: 11642	USGS NATIONAL WATER QUALITY LABORATORY POB 25046-MS 407-BLDG 95	Shipment Date : 24-Jul-2006
EPA Lab Code: CO00009	DENVER, CO 80225-0046 (303) 236-2000	Closing Date : 31-Aug-2006
	Director: MR. MERLE SHOCKEY	Score Date : 20-Sep-2006

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 295 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Purgeables								
trans-1,3-Dichloropropene EPA Code: N/A	ug/L	9513	USGS O-4127-96	33.3	32.3		20.6 - 44.1	Satisfactory 99 passed out of 101 reported results.
Tetrachloroethene ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	< 0.3	< 10.0 *			Satisfactory 104 passed out of 105 reported results.
Toluene ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	49.2	49.3 *	38.9 - 57.3	34.2 - 61.9	Satisfactory 114 passed out of 119 reported results.
Trichloroethene ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	68.2	66.3 *	49.5 - 79.1	42.1 - 86.5	Satisfactory 105 passed out of 107 reported results.
Total Xylenes EPA Code: N/A	ug/L	9513	USGS O-4127-96	308.9	260		157 - 364	Satisfactory 105 passed out of 115 reported results.
Vinyl chloride EPA Code: N/A	ug/L	9513	USGS O-4127-96	< 0.8	< 20.0 *			Satisfactory 98 passed out of 102 reported results.
Sample: Non Potable Water Residue								
Solids, Total EPA Code: N/A	mg/L	9502	USGS I-3750-85	467	466		422 - 510	Satisfactory 92 passed out of 101 reported results.
Sample: Non Potable Water Total Dissolved Solids								
Solids, Total Dissolved ** EPA Code: N/A	mg/L	9540	USGS I-1750-85	455	452 *	377 - 517	342 - 552	Satisfactory 100 passed out of 103 reported results.